



MS Autotune Report - Ultivo

Instrument Information

Model	G6465B Ultivo	Autotune Date	2026-09-05T12:49:08-07:00
Serial Number	SG1849Q103	SW/FW Version	2.4.467/8.1.38
Autotune Version	4.10.268	Last Autotune Date	2026-09-05T12:49:08-07:00
Ion Source	AJS ESI	Overall Result	Passed

Positive Results

Components

Ion Source Settings

Gas Temperature(°C)	300	Gas Flow(l/min)	7.0
Nebulizer(psi)	15	Sheath Gas Temperature(°C)	200
Sheath Gas Flow(l/min)	11.0	Capillary Voltage(V)	4000
Nozzle Voltage(V)	1500		

Optics Settings

Fragmentor(V)	120	Skimmer(V)	10
Cyclone Outer DC(V)	3	Cyclone Accelerator(V)	0
Cyclone Exit DC(V)	3	Cyclone Exit Lens(V)	2
Cyclone Inner RF(V)	400	Cyclone Outer RF(V)	350

Quad 1 Settings

MS1 Entrance Lens(V)	-50
----------------------	-----

MS1 Entrance Lens Dynamic Table

m/z	Setting
5.0	-28
50.0	-28
120.0	-29
320.0	-67
620.0	-124
920.0	-62
1220.0	0
1400.0	0

MS1 DC(V)	1	MS1 Heater(°C)	65
MS1 Exit Lens(V)	-50		

MS1 Exit Lens Dynamic Table

m/z	Setting
5.0	-36
50.0	-36
120.0	-36
320.0	-76
620.0	-137
920.0	-89
1220.0	-40
1400.0	-40

Collision Cell Settings

CC Entrance Lens(V)	0	CC Outer DC(V)	0
CAV(V)	7	CC Exit DC(V)	-7
CC Exit Lens(V)	-10	CC Inner RF(V)	400
CC Outer RF(V)	100		

Quad 2 Settings

MS2 Entrance Lens(V)	-100
----------------------	------

MS2 Entrance Lens Dynamic Table

m/z	Setting
5.0	-45
50.0	-45
120.0	-45
320.0	-54
620.0	-66
920.0	-72
1220.0	-78
1400.0	-78

MS2 DC(V)	-8	MS2 Heater(°C)	65
-----------	----	----------------	----

Detector Settings

Iris(V)	-100	HED(kV)	-10
EMV(V)	982		

Gain

Minimum Gain Factor	0.1	Maximum Gain Factor	10.0
Measured Gain Factor	1.0	Result	Passed

Mass Calibration Results

MS1 Peak Width Unit, Scan Speed Normal

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS1 Axis Gain	33.74	MS1 Axis Offset(amu)	0.779
---------------	-------	----------------------	-------


MS1 Axis Offset Dynamic Table

m/z	Setting
5.0	0.774
50.0	0.774
120.0	0.816
320.0	0.909
620.0	0.938
920.0	0.863
1220.0	0.776
1400.0	0.776

MS1 Width Gain(amu)	-41.3	MS1 Width Offset(amu)	-0.88
---------------------	-------	-----------------------	-------

MS1 Width Offset Dynamic Table

m/z	Setting
5.0	-0.87
50.0	-0.87
120.0	-0.94
320.0	-1.08
620.0	-1.07
920.0	-0.97
1220.0	-0.60
1400.0	-0.60



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
58.07	0.01	0.01	2.38E+5	98	Passed
118.09	0.00	0.01	6.12E+5	100	Passed
322.05	0.00	0.00	9.45E+4	100	Passed
622.03	0.01	0.01	1.52E+5	99	Passed
922.01	0.00	0.02	1.67E+5	99	Passed
1221.99	-0.01	0.00	1.63E+4	100	Passed

MS2 Peak Width Unit, Scan Speed Normal

Calibration Information

Score	98	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	10.46	MS2 Axis Offset(amu)	0.489
---------------	-------	----------------------	-------



MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	0.492
50.0	0.492
120.0	0.546
320.0	0.599
620.0	0.716
920.0	0.629
1220.0	0.558
1400.0	0.558

MS2 Width Gain(amu)	-40.2	MS2 Width Offset(amu)	-0.64
---------------------	-------	-----------------------	-------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	-0.68
50.0	-0.68
120.0	-0.69
320.0	-0.80
620.0	-0.65
920.0	-0.63
1220.0	-0.66
1400.0	-0.66



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
58.07	0.00	-0.01	3.04E+5	99	Passed
118.09	0.01	0.01	7.50E+5	99	Passed
322.05	-0.02	0.01	1.13E+5	97	Passed
622.03	0.00	0.02	2.78E+5	98	Passed
922.01	0.01	-0.02	2.14E+5	98	Passed
1221.99	0.00	0.01	8.43E+4	99	Passed

MS1 Peak Width Wide, Scan Speed Normal

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS1 Axis Gain	37.01	MS1 Axis Offset(amu)	0.868
---------------	-------	----------------------	-------


MS1 Axis Offset Dynamic Table

m/z	Setting
5.0	1.018
50.0	1.018
120.0	1.018
320.0	0.701
620.0	-0.004
920.0	-0.689
1220.0	-1.490
1400.0	-1.490

MS1 Width Gain(amu)	-44.9	MS1 Width Offset(amu)	-0.87
---------------------	-------	-----------------------	-------

MS1 Width Offset Dynamic Table

m/z	Setting
5.0	0.12
50.0	0.12
120.0	0.12
320.0	0.27
620.0	1.16
920.0	1.91
1220.0	2.81
1400.0	2.81



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
118.09	0.03	0.01	7.22E+5	98	Passed
322.05	0.00	0.02	1.13E+5	100	Passed
622.03	0.01	-0.01	2.33E+5	99	Passed
922.01	0.01	0.00	2.94E+5	100	Passed
1221.99	0.01	-0.04	2.92E+4	97	Passed

MS2 Peak Width Wide, Scan Speed Normal

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	9.58	MS2 Axis Offset(amu)	0.916
---------------	------	----------------------	-------


MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	0.900
50.0	0.900
120.0	0.900
320.0	1.092
620.0	1.317
920.0	1.418
1220.0	1.469
1400.0	1.469

MS2 Width Gain(amu)	-39.3	MS2 Width Offset(amu)	0.16
---------------------	-------	-----------------------	------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	0.11
50.0	0.11
120.0	0.11
320.0	-0.26
620.0	-0.12
920.0	-0.17
1220.0	-0.30
1400.0	-0.30



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
118.09	-0.02	-0.01	8.66E+5	99	Passed
322.05	-0.01	0.01	1.53E+5	99	Passed
622.03	-0.01	0.01	3.69E+5	100	Passed
922.01	0.05	-0.02	3.38E+5	95	Passed
1221.99	0.00	0.00	1.81E+5	100	Passed

MS1 Peak Width Widest, Scan Speed Normal

Calibration Information

Score	98	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS1 Axis Gain	34.06	MS1 Axis Offset(amu)	1.239
---------------	-------	----------------------	-------


MS1 Axis Offset Dynamic Table

m/z	Setting
5.0	2.053
50.0	2.053
120.0	2.053
320.0	2.158
620.0	1.873
920.0	1.827
1220.0	1.459
1400.0	1.459

MS1 Width Gain(amu)	-40.6	MS1 Width Offset(amu)	1.44
---------------------	-------	-----------------------	------

MS1 Width Offset Dynamic Table

m/z	Setting
5.0	1.95
50.0	1.95
120.0	1.95
320.0	1.63
620.0	1.80
920.0	1.74
1220.0	2.10
1400.0	2.10



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
118.09	-0.02	0.01	9.24E+5	99	Passed
322.05	-0.05	0.01	1.83E+5	98	Passed
622.03	-0.03	0.05	3.37E+5	98	Passed
922.01	0.03	-0.11	4.43E+5	96	Passed
1221.99	0.02	-0.03	4.80E+4	99	Passed

MS2 Peak Width Widest, Scan Speed Normal

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	9.58	MS2 Axis Offset(amu)	0.916
---------------	------	----------------------	-------


MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	1.871
50.0	1.871
120.0	1.871
320.0	2.084
620.0	2.119
920.0	2.112
1220.0	2.161
1400.0	2.161

MS2 Width Gain(amu)	-39.3	MS2 Width Offset(amu)	0.16
---------------------	-------	-----------------------	------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	2.10
50.0	2.10
120.0	2.10
320.0	1.88
620.0	2.07
920.0	2.11
1220.0	2.19
1400.0	2.19



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
118.09	0.03	-0.05	1.04E+6	99	Passed
322.05	-0.04	-0.02	2.31E+5	98	Passed
622.03	-0.01	-0.05	4.76E+5	99	Passed
922.01	0.04	-0.05	4.52E+5	98	Passed
1221.99	-0.02	0.03	2.68E+5	99	Passed

MS2 Scan Speed Fast

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	9.40	MS2 Axis Offset(amu)	1.899
---------------	------	----------------------	-------


MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	1.820
50.0	1.820
120.0	1.820
320.0	2.125
620.0	2.276
920.0	2.368
1220.0	2.516
1400.0	2.516

MS2 Width Gain(amu)	-41.6	MS2 Width Offset(amu)	0.76
---------------------	-------	-----------------------	------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	0.84
50.0	0.84
120.0	0.84
320.0	0.47
620.0	0.49
920.0	1.12
1220.0	0.83
1400.0	0.83



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
118.09	-0.02	0.01	9.22E+5	99	Passed
322.05	-0.03	0.03	1.64E+5	97	Passed
622.03	0.00	0.01	3.85E+5	100	Passed
922.01	-0.01	0.01	3.14E+5	100	Passed
1221.99	-0.02	0.00	1.14E+5	99	Passed

MS2 Scan Speed Ultra

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	10.37	MS2 Axis Offset(amu)	2.030
---------------	-------	----------------------	-------


MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	3.024
50.0	3.024
120.0	3.024
320.0	3.262
620.0	3.473
920.0	3.354
1220.0	3.330
1400.0	3.330

MS2 Width Gain(amu)	-42.1	MS2 Width Offset(amu)	2.18
---------------------	-------	-----------------------	------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	3.23
50.0	3.23
120.0	3.23
320.0	2.72
620.0	2.97
920.0	3.68
1220.0	3.78
1400.0	3.78



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
118.09	-0.01	0.03	9.90E+5	100	Passed
322.05	-0.02	0.03	2.00E+5	99	Passed
622.03	0.02	-0.04	4.49E+5	99	Passed
922.01	-0.03	-0.01	4.51E+5	100	Passed
1221.99	-0.01	0.01	2.44E+5	100	Passed

Negative Results

Components

Ion Source Settings

Gas Temperature(°C)	300	Gas Flow(l/min)	7.0
Nebulizer(psi)	15	Sheath Gas Temperature(°C)	200
Sheath Gas Flow(l/min)	11.0	Capillary Voltage(V)	4000
Nozzle Voltage(V)	1500		

Optics Settings

Fragmentor(V)	-90	Skimmer(V)	-10
Cyclone Outer DC(V)	-3	Cyclone Accelerator(V)	0
Cyclone Exit DC(V)	-3	Cyclone Exit Lens(V)	-2
Cyclone Inner RF(V)	400	Cyclone Outer RF(V)	350

Quad 1 Settings

MS1 Entrance Lens(V)	40
----------------------	----

MS1 Entrance Lens Dynamic Table

m/z	Setting
5.0	28
70.0	28
110.0	28
300.0	31
600.0	35
1030.0	39
1330.0	42
1400.0	42

MS1 DC(V)	-1	MS1 Heater(°C)	65
MS1 Exit Lens(V)	50		

MS1 Exit Lens Dynamic Table

m/z	Setting
5.0	27
70.0	27
110.0	27
300.0	32
600.0	42
1030.0	50
1330.0	56
1400.0	56

Collision Cell Settings

CC Entrance Lens(V)	0	CC Outer DC(V)	0
CAV(V)	-7	CC Exit DC(V)	7
CC Exit Lens(V)	10	CC Inner RF(V)	400
CC Outer RF(V)	100		

Quad 2 Settings

MS2 Entrance Lens(V)	100
----------------------	-----

MS2 Entrance Lens Dynamic Table

m/z	Setting
5.0	48
70.0	48
110.0	48
300.0	56
600.0	70
1030.0	77
1330.0	82
1400.0	82

MS2 DC(V)	8	MS2 Heater(°C)	65
-----------	---	----------------	----

Detector Settings

Iris(V)	100	HED(kV)	10
EMV(V)	1078		

Gain

Minimum Gain Factor	0.1	Maximum Gain Factor	10.0
Measured Gain Factor	1.0	Result	Passed

Mass Calibration Results**MS1 Peak Width Unit, Scan Speed Normal****Calibration Information**

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS1 Axis Gain	33.59	MS1 Axis Offset(amu)	0.675
---------------	-------	----------------------	-------


MS1 Axis Offset Dynamic Table

m/z	Setting
5.0	0.729
70.0	0.729
110.0	0.746
300.0	0.910
600.0	0.937
1030.0	0.950
1330.0	0.690
1400.0	0.690

MS1 Width Gain(amu)	-41.3	MS1 Width Offset(amu)	-0.97
---------------------	-------	-----------------------	-------

MS1 Width Offset Dynamic Table

m/z	Setting
5.0	-0.96
70.0	-0.96
110.0	-1.04
300.0	-1.02
600.0	-1.01
1030.0	-1.06
1330.0	-0.90
1400.0	-0.90



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
69.00	0.00	-0.01	2.79E+5	99	Passed
112.99	0.00	-0.01	6.48E+5	100	Passed
302.00	-0.01	-0.01	1.16E+5	99	Passed
601.98	0.00	0.02	1.10E+5	98	Passed
1033.99	0.00	-0.01	8.41E+5	99	Passed
1333.97	0.01	-0.01	1.52E+6	99	Passed

MS2 Peak Width Unit, Scan Speed Normal

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	10.67	MS2 Axis Offset(amu)	0.428
---------------	-------	----------------------	-------



MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	0.536
70.0	0.536
110.0	0.549
300.0	0.664
600.0	0.664
1030.0	0.568
1330.0	0.452
1400.0	0.452

MS2 Width Gain(amu)	-39.8	MS2 Width Offset(amu)	-0.71
---------------------	-------	-----------------------	-------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	-0.69
70.0	-0.69
110.0	-0.75
300.0	-0.88
600.0	-0.95
1030.0	-0.94
1330.0	-0.66
1400.0	-0.66



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
69.00	0.01	0.00	2.86E+5	98	Passed
112.99	0.01	0.01	8.10E+5	99	Passed
302.00	0.01	0.01	1.36E+5	98	Passed
601.98	0.00	0.00	1.37E+5	100	Passed
1033.99	0.00	0.01	1.45E+6	99	Passed
1333.97	0.00	-0.01	3.71E+6	100	Passed

MS1 Peak Width Wide, Scan Speed Normal

Calibration Information

Score	100	Result	Passed
-------	-----	--------	--------

Calibration Parameters

MS1 Axis Gain	32.49	MS1 Axis Offset(amu)	1.116
---------------	-------	----------------------	-------


MS1 Axis Offset Dynamic Table

m/z	Setting
5.0	1.155
70.0	1.155
110.0	1.155
300.0	1.420
600.0	1.534
1030.0	1.816
1330.0	1.652
1400.0	1.652

MS1 Width Gain(amu)	-40.9	MS1 Width Offset(amu)	-0.18
---------------------	-------	-----------------------	-------

MS1 Width Offset Dynamic Table

m/z	Setting
5.0	-0.15
70.0	-0.15
110.0	-0.15
300.0	-0.29
600.0	-0.40
1030.0	-0.39
1330.0	-0.44
1400.0	-0.44



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
112.99	0.01	0.01	8.17E+5	100	Passed
302.00	0.01	-0.02	1.31E+5	100	Passed
601.98	0.00	0.01	1.39E+5	100	Passed
1033.99	0.00	0.00	1.49E+6	100	Passed
1333.97	0.01	0.01	2.64E+6	100	Passed

MS2 Peak Width Wide, Scan Speed Normal

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	9.51	MS2 Axis Offset(amu)	0.600
---------------	------	----------------------	-------


MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	0.976
70.0	0.976
110.0	0.976
300.0	1.113
600.0	1.302
1030.0	1.577
1330.0	1.618
1400.0	1.618

MS2 Width Gain(amu)	-40.2	MS2 Width Offset(amu)	-0.70
---------------------	-------	-----------------------	-------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	0.14
70.0	0.14
110.0	0.14
300.0	0.05
600.0	0.00
1030.0	0.10
1330.0	0.37
1400.0	0.37



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
112.99	0.00	-0.01	9.81E+5	100	Passed
302.00	0.00	0.02	1.56E+5	100	Passed
601.98	-0.01	0.02	1.66E+5	99	Passed
1033.99	0.01	-0.04	2.38E+6	98	Passed
1333.97	-0.01	-0.03	5.66E+6	99	Passed

MS1 Peak Width Widest, Scan Speed Normal

Calibration Information

Score	98	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS1 Axis Gain	32.49	MS1 Axis Offset(amu)	1.116
---------------	-------	----------------------	-------


MS1 Axis Offset Dynamic Table

m/z	Setting
5.0	2.118
70.0	2.118
110.0	2.118
300.0	2.252
600.0	2.318
1030.0	2.534
1330.0	2.407
1400.0	2.407

MS1 Width Gain(amu)	-40.9	MS1 Width Offset(amu)	-0.18
---------------------	-------	-----------------------	-------

MS1 Width Offset Dynamic Table

m/z	Setting
5.0	1.92
70.0	1.92
110.0	1.92
300.0	1.78
600.0	1.70
1030.0	1.92
1330.0	2.06
1400.0	2.06



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
112.99	0.01	-0.03	9.90E+5	100	Passed
302.00	0.00	0.08	1.57E+5	98	Passed
601.98	-0.03	-0.11	1.83E+5	96	Passed
1033.99	-0.03	-0.01	2.22E+6	99	Passed
1333.97	-0.05	0.01	4.52E+6	97	Passed

MS2 Peak Width Widest, Scan Speed Normal

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	9.51	MS2 Axis Offset(amu)	0.600
---------------	------	----------------------	-------


MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	1.875
70.0	1.875
110.0	1.875
300.0	1.953
600.0	2.153
1030.0	2.301
1330.0	2.347
1400.0	2.347

MS2 Width Gain(amu)	-40.2	MS2 Width Offset(amu)	-0.70
---------------------	-------	-----------------------	-------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	2.16
70.0	2.16
110.0	2.16
300.0	2.15
600.0	2.28
1030.0	2.38
1330.0	2.75
1400.0	2.75



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
112.99	0.01	-0.05	1.13E+6	99	Passed
302.00	0.00	-0.01	1.69E+5	100	Passed
601.98	-0.02	0.06	1.98E+5	99	Passed
1033.99	0.01	-0.04	3.37E+6	99	Passed
1333.97	0.02	-0.05	8.74E+6	99	Passed

MS2 Scan Speed Fast

Calibration Information

Score	98	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	9.51	MS2 Axis Offset(amu)	1.781
---------------	------	----------------------	-------


MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	1.927
70.0	1.927
110.0	1.927
300.0	2.141
600.0	2.202
1030.0	2.461
1330.0	2.457
1400.0	2.457

MS2 Width Gain(amu)	-42.3	MS2 Width Offset(amu)	0.94
---------------------	-------	-----------------------	------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	0.85
70.0	0.85
110.0	0.85
300.0	0.69
600.0	0.40
1030.0	0.87
1330.0	1.13
1400.0	1.13



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
112.99	0.02	-0.03	1.21E+6	98	Passed
302.00	-0.02	0.02	1.74E+5	98	Passed
601.98	0.01	0.00	1.78E+5	100	Passed
1033.99	-0.02	0.04	1.89E+6	96	Passed
1333.97	0.01	-0.01	3.82E+6	100	Passed

MS2 Scan Speed Ultra

Calibration Information

Score	99	Result	Passed
-------	----	--------	--------

Calibration Parameters

MS2 Axis Gain	10.18	MS2 Axis Offset(amu)	2.060
---------------	-------	----------------------	-------


MS2 Axis Offset Dynamic Table

m/z	Setting
5.0	3.248
70.0	3.248
110.0	3.248
300.0	3.185
600.0	3.414
1030.0	3.796
1330.0	3.571
1400.0	3.571

MS2 Width Gain(amu)	-43.0	MS2 Width Offset(amu)	2.36
---------------------	-------	-----------------------	------

MS2 Width Offset Dynamic Table

m/z	Setting
5.0	3.43
70.0	3.43
110.0	3.43
300.0	2.90
600.0	2.81
1030.0	3.65
1330.0	4.00
1400.0	4.00



Mass Calibration Results

Theoretical (m/z)	Mass Delta(m/z)	Width Delta(m/z)	Abundance	Score	Result
112.99	0.02	0.05	1.32E+6	99	Passed
302.00	-0.03	-0.03	2.04E+5	99	Passed
601.98	0.03	0.02	2.19E+5	99	Passed
1033.99	0.06	-0.03	3.39E+6	98	Passed
1333.97	0.01	-0.05	7.27E+6	99	Passed

Vacuum And Temperature

Rough Vac(Torr) 1.56E+0
Turbo Speed(%) 99.9
MS2 Heater(°C) 65

High Vac(Torr) 5.90E-5
MS1 Heater(°C) 65